

# MRC Holland Support

Help Centre > digitalMLPA & Coffalyser digitalMLPA > digitalMLPA Technique > Do mismatches between probes and their target sequences influence (digital)MLPA results?

---

## Do mismatches between probes and their target sequences influence (digital)MLPA results?

This article was retrieved from MRC Holland Support (support.mrcholland.com) on Monday, 30th June 2025.

Probes can be affected by mismatches between the probe sequence and the sample DNA. The presence of one or more mismatches can lead to a reduced probe signal, which can mimic a deletion. As this is also the case for non-pathogenic variants, such as SNVs, it is very important to confirm single-probe deletions. Read more about confirming results in [this article](#).

## Background

Mismatches can have different effects depending on their position in the probe's target site.

- Mismatches at the ligation site always disrupt the ligation of the two probe oligonucleotides to some extent, depending on the type of mismatch. This is what enables the detection of known SNVs with conventional MLPA and digitalMLPA. See [this article](#) for more information.
- Mismatches near the ligation site (within ~5 nt) may also decrease the ligation efficiency, which can result in a decreased signal.
- Mismatches further from the ligation site may sometimes have an effect if they destabilize the probe's hybridisation to the sample DNA. This is less likely to be an issue for digitalMLPA probes, as these probes tend to have longer hybridisation sequences than conventional MLPA probes.

When we design probes, we try to avoid frequently occurring variants retrieved from public databases and customer feedback. However, new variants are continuously being discovered, and it is often not possible to avoid all known (rare) variants. For this reason, the possibility of a mismatch should always be considered for a single-probe deletion.

### Note

Customer feedback is a very important source of information about the performance of our products in practice. If you find a variant with an effect on one of our probes, please [contact us](#) – especially if the variant occurs frequently. We may be able to use this information to improve the probe.

### Tags

digitalMLPA

MLPA

### Related Content

- [How can I confirm my \(digital\)MLPA results?](#)
- [Is it possible to detect SNVs with \(digital\)MLPA?](#)

#### Disclaimer

The information provided in this material is correct for the majority of our products. For certain applications, the instructions for use may differ. In the event of conflicting information, the relevant instructions for use take precedence.