

# MRC Holland Support

Support > Help Centre > MLPA & Coffalyser.Net > Raw Data Interpretation & Troubleshooting > [Is it possible to compare results from different \(digital\)MLPA experiments/runs?](#)

---

## Is it possible to compare results from different (digital)MLPA experiments/runs?

This article was retrieved from [support.mrcholland.com](https://support.mrcholland.com) on Wednesday, 24th April 2024.

It is not possible to combine raw data or other intermediates from the data analysis from different conventional MLPA or digitalMLPA experiments into a single analysis. It is possible to combine data from two sequencing runs on the same digitalMLPA products to obtain sufficient reads for data analysis.

## Background

Conventional MLPA and digitalMLPA are relative techniques. All steps in the experimental procedure can introduce [variation](#). Therefore, test samples and reference samples must be part of the same experiment, using the same master mixes. The final probe ratios that are obtained *after* normalisation can be compared, as the normalisation will have removed most of the experimental variation.

In digitalMLPA experiments, it is also important that test and reference samples are part of the same sequencing run in the same flow cell to limit variability. There is only one scenario in which data from two different sequencing runs can be combined in a single analysis, and this is when the *entire* run is repeated to obtain more reads. In this scenario both runs can be combined, as test and reference samples have still been treated equally.

Tags

digitalMLPA

MLPA

### Related Pages

---

- [Which factors influence variability of \(digital\)MLPA probes?](#)

### Disclaimer

---

*The information provided in this material is correct for the majority of our products. However, for certain applications, the instructions for use may differ. In the event of*

*conflicting information, the relevant instructions for use take precedence.*