

MRC Holland Support

Support > Help Centre > Melt Assays > What are the sample requirements for the MC002 SMA Newborn Screen?

What are the sample requirements for the MC002 SMA Newborn Screen?

This article was retrieved from support.mrcholland.com on Friday, 9th May 2025.

[SALSA® MC002 SMA Newborn Screen](#) can be used with crude DNA extracts from dried blood spot (DBS) cards. Crude extracts can be prepared from 1.5 or 3.2 mm punches of DBS cards based on untreated Whatman 903 paper. DNA must be extracted via one of the protocols given in the [instructions for use](#).

Only DBS cards based on untreated Whatman 903 paper have been validated. For validation of other cards, please [contact us](#). Cards impregnated with chemicals, such as FTA cards, are not suitable.

Note

Please see the [instructions for use](#) for the most up-to-date information.

Tags

Melt Assay

Related Pages

- [Is a specific thermocycler required for the MC002 SMA Newborn Screen?](#)
- [How can I confirm results obtained with the MC002 SMA Newborn Screen?](#)

Disclaimer

The information provided in this material is correct for the majority of our products. However, for certain applications, the instructions for use may differ. In the event of conflicting information, the relevant instructions for use take precedence.